



## **Product Information** **EZ-LINE Cell Culture Dishes, Treated**

### **Product information for SP31146, SP31171, SP31212:**

#### **Description**

Cell culture dishes for tissue culture work. All dishes are produced with optically clear, high quality crystalline polymers. Growth surface is treated for optimal cell attachment and growth and designed for reduced cross-contamination.



#### **Features:**

- Sterilized with gamma irradiation
- Non-pyrogenic
- TC treated
- High strength

Item No.	Description	Package
SP31146	EZ-LINE 35x10mm Cell Culture Dish, TC Treated, sterile, 10 per pack, 500 per case	500/case
SP31171	EZ-LINE 60x15mm Cell Culture Dish, TC Treated, sterile, 10 per pack, 500 per case	500/case
SP31212	EZ-LINE 90x20mm Cell Culture Dish, TC Treated, Sterile, 10 per pack, 200 per case	200/case

### **Technical Specifications**

Item No.	SP31146	SP31171	SP31212
Quantity per case	500/case	500/case	200/case
Packaging	10 Pack	10 Pack	10 Pack
Volume	3mL	5mL	12.5mL
Growth Area	9.40cm <sup>2</sup>	21.5cm <sup>2</sup>	57.0cm <sup>2</sup>
Color	Clear	Clear	Clear
Plastic materials	Polystyrene	Polystyrene	Polystyrene
Sterile	Yes	Yes	Yes
TC Treated	Yes	Yes	Yes

#### **Storage**

Room temperature.

QF 24 Rev 0



### Certificate of Compliance

**Product Description:** Cell culture products  
**Expiration Date:** 5 years from manufacture date

**Visual Attribute:** PASS  
**Functional Test:** PASS  
**Sterilization:** Sterilization was performed by exposed gamma irradiation according to ISO 11137 based on the *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization <SAL 10-6>*  
**Cytotoxicity Test:** Cytotoxicity test is conducted to qualify all material resins using USP class VI<87> and/or ISO 10993 standards for cytotoxicity and have been shown to be non-toxic  
**Pyrogens Test:** Non-pyrogenic tested according to the principles of the LAL test described in USP<85>/ 0.5EU/ml  
**RNase/DNase Test:** This product has been tested and is free of any detectable RNase/DNase contamination

- ✓ Bio Basic Inc. hereby certifies that the products identified above have been manufactured & inspected in compliance with products quality specifications and requirements as documented in our ISO 9001:2008 Quality Management System.